

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL



**Remarque :** Selon le dispositif médical (DM) concerne, ce dossier concernera une référence, un type ou une famille de DM

<b>1. Renseignements administratifs concernant l'entreprise</b>		<b>Date de mise à jour : 03/11/2016</b> <b>Date d'édition : 03/11/2016</b>
1.1	<b>Nom :</b> Laboratoires EUROMEDIS	
1.2	<b>Adresse complète :</b> Z.I. de la Tuilerie 60290 NEUILLY -SOUS- CLERMONT	<b>Tel:</b> +33 344738360 <b>Fax :</b> +33 344735732 <b>e-mail :</b> euromedis@euromedis.fr <b>Site internet :</b> www.euromedis.fr
1.3	<b>Coordonnées du correspondant matériovigilance :</b> M. Eddie ZERBIB Responsable qualité	<b>Tel :</b> +33 344738360 <b>Fax :</b> +33 344735732 <b>e-mail :</b> eddie.zerbib@euromedis.fr

<b>2. Informations sur dispositif ou équipement</b>			
2.1	<b>Dénomination commune :</b> selon la nomenclature d' Europharmat®		
2.2	<b>Dénomination commerciale :</b> FLEXISKIN® GANT EXAMEN NITRILE BLEU COBALT SANS POUDRE 240 mm		
2.3	<b>Code nomenclature :</b> 11882		
2.4	<b>Code LPPR* (ex TIPS si applicable) :</b>		
2.5	<table border="1"> <tr> <td><b>Classe du DM :</b> Classe I <b>Directive de l'UE applicable :</b> 93/42/CE <b>Selon Annexe n° :</b> IX chapitre 3 et VII chapitre 3 <b>Déclaration CE de conformité</b></td> <td><b>Catégorie de l'EPI :</b> Catégorie III <b>Directive de l'UE applicable :</b> 89/686/CE <b>Selon Annexe n° :</b> chapitre II article 11 <b>Certificat CE0465 du CIMAC</b></td> </tr> </table> <p><b>Date de première mise sur le marché dans l'UE :</b> 08/2016</p> <p><b>Fabricant du DM :</b> Laboratoires EUROMEDIS <b>Certificat applicable à l'entreprise fabricante :</b> ISO 9001 : 2008 et ISO 13485 : 2003 <b>Organisme certificateur :</b> LNE/G-MED</p> <p><b>Normes applicable au dispositif médical :</b></p> <ul style="list-style-type: none"> <li>- Directive 93/42 CE pour les dispositifs médicaux EN 455-1/2/3/4- Gants médicaux non réutilisables</li> <li>- Directive 89/686 CE Pour les équipements de protection individuels EN 374-1/2/3- Gants de protection contre les produits chimiques et les micro-organismes EN 420- Gants de protections exigences générales et méthodes d'essais EN 388 - Gants de protection contre les risques mécaniques</li> </ul>	<b>Classe du DM :</b> Classe I <b>Directive de l'UE applicable :</b> 93/42/CE <b>Selon Annexe n° :</b> IX chapitre 3 et VII chapitre 3 <b>Déclaration CE de conformité</b>	<b>Catégorie de l'EPI :</b> Catégorie III <b>Directive de l'UE applicable :</b> 89/686/CE <b>Selon Annexe n° :</b> chapitre II article 11 <b>Certificat CE0465 du CIMAC</b>
<b>Classe du DM :</b> Classe I <b>Directive de l'UE applicable :</b> 93/42/CE <b>Selon Annexe n° :</b> IX chapitre 3 et VII chapitre 3 <b>Déclaration CE de conformité</b>	<b>Catégorie de l'EPI :</b> Catégorie III <b>Directive de l'UE applicable :</b> 89/686/CE <b>Selon Annexe n° :</b> chapitre II article 11 <b>Certificat CE0465 du CIMAC</b>		

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL

2.6

**Descriptif du dispositif (avec photo, schéma, dimensions, volume, ...):** peut être relié au point 8 : selon fiche technique.

Gant nitrile bleu-cobalt offrant une très bonne résistance aux agressions des produits chimiques et une souplesse étonnante permettant un toucher de grande sensibilité

Gant d'examen ambidextre existe en 5 tailles T 5/6 - T 6/7 - T 7/8 - T 8/9 - T 9/10

Usage Unique : Oui

Couleur : Bleu- Violet

Texture : extrémité distale rugueuse

Forme : Ambidextre

Bord : Roulé

Alimentaire : Oui sauf pour les amines

Code couleur sur le packaging : Oui, voir tableau des dimensions.

Origine : Asie du Sud Est

Trousse : Non

Dimension du dispositif : (Voir annexe 1)

Taille		Longueur en mm Mini	Périmètre de la paume en mm	Epaisseur en mm ( $\pm 0.01$ )		
				Manchette	Paume	Doigt
T 5/6	XS	240	70-79	0.05	0.05	0.09
T 6/7	S	240	80-89	0.05	0.05	0.09
T 7/8	M	240	90-99	0.05	0.05	0.09
T 8/9	L	240	100-109	0.05	0.05	0.09
T 9/10	XL	240	110-119	0.05	0.05	0.09

2.7

**Références Catalogue :**

Pour chaque référence préciser :

**REFERENCE :**

Conditionnement / emballages

**UCD** (Unité de Commande) : La boîte

**CDT** (Multiple de l'UCD) : Quantité par carton

**QML** (Quantité minimale de livraison) : Le carton

Spécification du produit	Référence	Unités / boîte	Boîtes / carton	Unités / carton
T 5/6	127595	100	10	1000
T 6/7	127596	100	10	1000
T 7/8	127597	100	10	1000
T 8/9	127598	100	10	1000
T 9/10	127599	100	10	1000

2.8

**Composition du dispositif et Accessoires :**

Latex : Non

Agent de vulcanisation : Oui

Présence de DEHP: Non

Produit d'origine animale ou biologique : Non

- Nitrile
- Phtalocyanine blue
- Oxyde de Zinc
- Sulphure
- Dioxyde de titane
- Zinc dibutyldithiocarbamate
- Hydroxyde de potassium

**Dispositifs et accessoires associés à lister.** NA

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL

### Caractéristiques de la référence :

Norme	Essai	Résultats
EN 455-1	Etanchéité	Niveau inspection 1 : AQL=1.5
EN 455-2	Force minimale à la rupture	
	- Avant vieillissement accéléré	≥ 6 N
	- Après vieillissement accéléré:	≥ 6 N
EN 455-3	Taux de poudre résiduel	<2 mg/gant
EN 455-3	Taux de protéine	Indétectable
EN 455-4	Détermination de la durée de conservation	5 ans
EN 374-1	Terminologie	Conforme
EN 374-2	Essai de fuite à l'eau	Conforme
EN 374-2	Essai de fuite à l'air	Conforme
EN 374-3	(L) Acide sulfurique 96%	> 30 min index 2
	(K) Hydroxyde de sodium 40%:	> 90 min index 3
	(G) Diethylamine	> 30 min index 2
EN 420	Taille et dimension	OK
EN 388	Résistance à l'abrasion	Niveau de performance = 0
	Résistance à la coupure	Niveau de performance = 0
	Résistance au déchirement	Niveau de performance = 0
	Résistance à la perforation	Niveau de performance = 0
ISO 10993-1	Cytotoxicité	Conforme
	Sensibilisation	Conforme
	Irritation	Conforme
ASTM D 6978-05	Test des drogues de chimiothérapie	Testé voir page5
ASTM F 1671-07	Penetration viral et bactériologique	Testé voir page10
<b>Alimentarité</b>		
Directive 2002/72/CE du 6/08/2002	-Essai de migration globale	voir page 11, la déclaration de conformité à la réglementation relative aux matériaux des matériels et équipements au contact des denrées alimentaires

Résultats du rapport PN109972B ( page 5) du test de perméation suivant l'ASTM D 6978 ( EN 374-3)

<b>CYTOTOXIQUE</b>	Indice de performance à la perméation
<b>Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)</b>	<b>0</b>
<b>Cisplatine, 1.0 mg/ml (1,000 ppm)</b>	<b>5</b>
<b>Cyclophosphamide (Cytosan), 20.0 mg/ml (20,000ppm)</b>	<b>5</b>
<b>Dacarbazine (DTIC), 10.0 mg/ml(10,000 ppm)</b>	<b>5</b>
<b>Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)</b>	<b>5</b>
<b>Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)</b>	<b>5</b>
<b>Fluorouracil, 50.0 mg/ml (50,000 ppm)</b>	<b>5</b>
<b>Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)</b>	<b>5</b>
<b>Thiotepa, 10.0 mg/ml (10,000 ppm)</b>	<b>2</b>

Tableau 1 — Indices de performance à la perméation

Temps de passage mesuré (min)	Indice de performance à la perméation
≤ 10	1
> 10	2
> 30	3
> 60	4
> 120	5
> 240	6
> 480	6

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL

<b>2.9</b>	<b>Domaine - Indications :</b> Domaine d'utilisation: <i>Médical et Industriel</i> Indications: <i>Gant examen résistant aux agents chimiques</i>
<b>3.</b>	<b>Procédé de stérilisation :</b>
	<b>DM stérile :</b> <i>Non</i>
<b>4.</b>	<b>Conditions de conservation et de stockage</b>
	Conditions normales de conservation & de stockage : <i>ne doit pas être exposé à l'humidité et au soleil</i> Précautions particulières : <i>Usage unique</i> Durée de la validité du produit : <i>5 ans</i> Présence d'indicateurs de température s'il y a lieu : <i>NA</i>
<b>5.</b>	<b>Sécurité d'utilisation</b>
	<b>Sécurité technique :</b> <i>Conforme aux normes EN 455-1/2/3/4, EN 374-1/2/3 et EN 420</i>
	<b>Sécurité biologique:</b> <i>NA</i>
<b>6.</b>	<b>Conseils d'utilisation</b>
<b>6.1</b>	<b>Mode d'emploi :</b> <i>NA</i>
<b>6.2</b>	<b>Indications :</b> (destination marquage CE) <i>Examen Médical - Protection du patient et de l'utilisateur lors de soins</i>
<b>6.3</b>	<b>Précautions d'emploi :</b> <i>Ne pas ouvrir avec un objet coupant</i>
<b>6.4</b>	<b>Contre- Indications :</b> <i>NA</i>
<b>7.</b>	<b>Informations complémentaires sur le produit</b>
	<b><u>Bibliographie, rapport d'essais cliniques, ou d'études pharmaco-économiques, amélioration du service rendu : recommandations particulières d'utilisation (restrictions de prise en charge, plateau technique, qualification de l'opérateur, etc.) ... :</u></b> <i>Toutes les informations se trouvent dans le dossier technique</i>
<b>8.</b>	<b>Liste des annexes au dossier (s'il y a lieu)</b>
	- <i>Etiquetage et étiquette de traçabilité (le cas échéant)</i> <i>Etiquetage conforme à la directive</i>

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL

Testing. Development. Problem Solving.



July 17, 2013

### • TEST REPORT •

PN 109972B

### CHEMICAL ANALYTICAL SERVICES

Prepared For:

Prepared By

  
Jeffrey L. Keller  
Assistant Manager  
Pharmaceutical Services

Approved By

  
Ana C. Barbur, M.S.  
Manager  
Chemical, Microbiological, & Pharmaceutical Services



An A2LA Accredited Testing Laboratory -- Certificate Numbers 255.01 & 255.02  
ISO 9001:2008 Registered

ISO 9001:2008  
Registered

... offers and reports are for the exclusive use of the clients to whom they are addressed and shall not be reproduced, except in full, without the written permission of Akron Rubber Development Laboratory, Inc. (ARDL). The information contained herein applies to the specific product, process or process control of evaluation. The accuracy of any kind of test is certified as required. The liability of ARDL, Inc. shall be limited to the amount of compensation paid for services. ARDL, Inc. is accredited by A2LA for the test methods listed on the attached scope.

[www.ardl.com](http://www.ardl.com)

2887 Gilchrist Rd | Akron, Ohio 44305 | [answers@ardl.com](mailto:answers@ardl.com)  
Toll Free (800) 830-ARDL | Worldwide (330) 794-6600 | Fax (330) 794-6610

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL



Testing, Development, Problem Solving.

July 17, 2013

Page 1 of 3 – PN 109972B

**SUBJECT:** Permeation testing per ASTM D 6978-05 on sample submitted by the above company.

**RECEIVED:** Glove sample identified as Powder Free Nitrile Examination Gloves; Lot# S2123286 (32BP), Size Medium.

**TESTING CHEMOTHERAPY DRUGS:**

Table 1. List of the Testing Chemotherapy Drugs, Sources, and Expiration Dates

TESTING CHEMOTHERAPY DRUGS	DRUG SOURCE
Carmustine (BCNU)	Bristol-Myers, Lot# 2C7009A, Expiration 07/2015
Cisplatin	Pfizer, Lot# 7800808, Expiration 08/2013
Cyclophosphamide (Cytosan)	Sigma, Lot# SLBC068V, Expiration 03/2014
Dacarbazine (DTIC)	APP, Lot# 6103390, Expiration 02/2014
Doxorubicin Hydrochloride	USP, Lot# LOK258, Expiration 06/2014
Etoposide (Tosopar)	Teva, Lot# 31314884B, Expiration 02/2015
Fluorouracil	APP, Lot# 6104355, Expiration 01/2014
Paclitaxel (Taxol)	Hospira, Lot# Z086885AA, Expiration 04/2014
Thiotepa	Sigma Aldrich, Lot# SLBD4239V, Expiration 10/2014

**COLLECTION MEDIA:**

The collection media, which were selected, are listed in Table 2.

Table 2. Collection Media for Testing Chemotherapy Drugs

TEST DRUG AND CONCENTRATION	COLLECTION MEDIUM
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	10% Ethanol Aqueous Solution
Cisplatin, 1.0 mg/ml (1,000 ppm)	Distilled Water
Cyclophosphamide (Cytosan), 20 mg/ml (20,000 ppm)	Distilled Water
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	Distilled Water
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	Distilled Water
Etoposide (Tosopar), 20.0 mg/ml (20,000 ppm)	Distilled Water
Fluorouracil, 50.0 mg/ml (50,000 ppm)	8.20 pH Sodium Hydroxide Solution
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	30% Methanol Aqueous Solution
Thiotepa, 10.0 mg/ml (10,000 ppm)	Distilled Water

www.ardl.com

2887 Gilchrist Rd. | Akron, Ohio 44305 | answers@ardl.com  
 Toll Free (800) 830-ARDL | Worldwide (330) 794-8800 | Fax (330) 794-6610

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL

Page 2 of 3 – PN 1099728

### TESTING CONDITIONS:

Standard Test Method Used:	ASTM D 6978-06
Analytical Method:	UV/VIS Spectrometry
Testing Temperature:	35.0°C ± 2.0
Collection System:	Closed Loop
Specimen Area Exposed:	5.067 cm <sup>2</sup>
Selected Data Points:	25/test
Number of Specimens Tested:	3/test
Location Sampled From:	Cuff area
Comments/Other Conditions:	Magnetic stir bar was used in the sampling chamber

### DETECTION METHOD OF CHEMICAL PERMEATION: UV/VIS ABSORPTION SPECTROMETRY:

Instrument: Perkin Elmer UV/VIS Spectrometer Lambda 25

UV/VIS Absorption Spectrometry was used to measure the absorbance of test chemicals, which permeated through the specimens into the collection medium. The collection medium was circulated in a closed loop at 11 ml/minute of flow rate through the testing period. Data collection was performed according to the programmed schedule by means of UV Winlab software from the Perkin Elmer Corporation. The list of the characteristic wavelengths is shown below.

Table 3. Characteristic Wavelengths used in UV/VIS Absorption Spectrometry

TESTING CHEMOTHERAPY DRUGS	WAVELENGTH (nm)
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	229
Cisplatin, 1.0 mg/ml (1,000 ppm)	199
Cyclophosphamide (Cytosar), 20 mg/ml (20,000 ppm)	200
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	320
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	232
Etoposide (Tosopar), 20.0 mg/ml (20,000 ppm)	205
Fluorouracil, 60.0 mg/ml (60,000 ppm)	269
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	231
Thiotepa, 10.0 mg/ml (10,000 ppm)	199

### SAMPLE CHARACTERISTICS:

Table 4. Thickness characteristics for the tested specimens: Powder Free Nitrile Examination Gloves, Lot# S2123266 (32BP), Size Medium.

Testing Chemotherapy Drugs	Thickness (mm)			Average (mm)	Weight/Unit Area (g/m <sup>2</sup> )
	Sample 1	Sample 2	Sample 3		
Carmustine (BCNU)	0.060	0.065	0.055	0.060	58.4
Cisplatin	0.058	0.062	0.057	0.059	58.4
Cyclophosphamide (Cytosar)	0.070	0.060	0.060	0.063	58.4
Dacarbazine (DTIC)	0.061	0.063	0.059	0.061	58.4
Doxorubicin Hydrochloride	0.068	0.063	0.057	0.062	58.4
Etoposide (Tosopar)	0.064	0.067	0.058	0.060	58.4
Fluorouracil	0.060	0.059	0.058	0.059	58.4
Paclitaxel (Taxol)	0.063	0.063	0.058	0.061	58.4
Thiotepa	0.057	0.060	0.055	0.061	58.4

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL

Page 3 of 3 – PN 109972B

**RESULTS:**

Table 5. Permeation Test Results on Powder Free Nitrile Examination Gloves, Lot# S2123286 (32BP), Size Medium.

TEST CHEMOTHERAPY DRUG AND CONCENTRATION	MINIMUM BREAKTHROUGH DETECTION TIME (Specimen 1/2/3) (Minutes)	STEADY STATE PERM. RATE (Specimen 1/2/3) (µg/cm <sup>2</sup> /minute)	OTHER OBSERVATIONS
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	1.0 (1.2, 2.1, 1.0)	2.1 (2.3, 2.0, 2.0)	Moderate swelling and no degradation
Cisplatin, 1.0 mg/ml (1,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Cyclophosphamide (Cytosan), 20 mg/ml (20,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Fluorouracil, 50.0 mg/ml (50,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	No breakthrough up to 240 min.	N/A	Moderate swelling and no degradation
Thiotepa, 10.0 mg/ml (10,000 ppm)	31.3 (45.5, 31.3, 46.4)	1.1 (1.4, 0.6, 1.2)	Slight swelling and no degradation



Tiffany L. Heller  
Assistant Manager  
Pharmaceutical Services  
AKRON RUBBER DEVELOPMENT LABORATORY, INC



Ana C. Barbur, M.S.  
Manager  
Chemical, Microbiological and Pharmaceutical Services



# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL

ASTM F 1671-07 - Penetration viral et bactériologique



**LEMBAGA GETAH MALAYSIA**  
*Malaysian Rubber Board*

Ibu Pejabat Lembaga : Tingkat 18, Bangunan Getah Asli (Menara),  
148, Jalan Ampang, 50450 Kuala Lumpur, Malaysia.  
Peti Surat 10150, 50908 Kuala Lumpur, Malaysia. Tel : +60(0)3 92062000 Fax : +60(0)3 21634492  
E-mail : general@lgm.gov.my Web-site : http://www.lgm.gov.my



PAGE: 1 of 2

### TEST REPORT

**REPORT NUMBER** : USTL/2014/05/0001

**DATE**: 22/6/14

**SUBJECT** : VIRUS PENETRATION TEST

**RECEIVED ON** : 9 May 2014

These results have been obtained on sample(s) submitted to us.

**Condition of samples** : Unused gloves with no wear or abrasion

**Expected** : No penetration of viral solution from inside the glove; any  
penetration above 10 pfu is considered failed. Expected recovery;  
100 ± 2%

**Test objective** : To determine that the viral solution did not penetrate > 10 pfu  
within the test period (viral leak < 10 pfu)

# Dossier d'information type Euro Pharmat

## DISPOSITIF MEDICAL



### LEMBAGA GETAH MALAYSIA Malaysian Rubber Board

(bu Pejabat Lembaga : Tingkat 18, Bangunan Getah Asli (Menara),  
14B, Jalan Ampang, 50450 Kuala Lumpur, Malaysia.  
Peta Surai 10150, 50908 Kuala Lumpur, Malaysia. Tel : +60(0)3 92062000 Fnx : +60(0)3 21634492  
E-mail : general@lgm.gov.my Web-Site : http://www.lgm.gov.my



PAGE: 2 of 2

## TEST REPORT

**REPORT NUMBER** : USTL/2014/05/0001 **DATE:** 22/6/14

**SAMPLE DESRIPTION** : **Powder Free Nitrile Examination Gloves**

**Lot No:** 8418291132

**STANDARD TEST METHOD** : *ASTM F1671-07*, Phi-X174 Bacteriophage Penetration Test

**DATE OF TESTING** : 10 June 2014

No.	Sample Id	Plague	Requirement	Status
		Forming Unit	Virus Leak (Pfu)	
1	Powder Free Nitrile Examination Gloves Lot No: 8418291132	NP	<10	PASS

Note

NP : No plaque formed  
VSF\* : Test results are acceptable if VSF > 0.8

Disclaimer

Test is performed to required specification (s) of the said standard (where applicable). Results reflect data obtained and/or observed from the samples provided for testing only. Results do not reflect shipment prior to the stated lot numbers, or condition of future shipment, nor does it reflect the quality of future production and manufacturing. Our organization is not liable for any mis-used of data or information

Yours Sincerely;

(TAJUL ANUAR YAAKOB)  
Research Officer  
47000 Sg Buloh  
Selangor, Malaysia

# Dossier d'information type Euro Pharmat

DISPOSITIF MEDICAL

**A.N.C.I. Servizi S.r.l.**

a socio unico

Sede legale e amministrativa

20149 MILANO

Via Monte Rosa, 21

Tel. 02.438291

Fax 02.48005833

Cap. Soc.: € 10.400 i.v.

C.F./P.I.: 07199040150

Reg. Imprese n° 229059

Trib. di Milano

R.E.A. n° 1147818



**CIMAC**

Centro Italiano  
Materiali di Applicazione  
Calzaturiera



PRD N° 0171 B

Membro degli Accordi di Mutuo Riconoscimento  
EA, IAF e ILAC

Signatory of EA, IAF and ILAC  
Mutual Recognition Agreements

Sede operativa: 27029 VIGEVANO (PV) - C.so G. Brodolini, 19 - Tel. 0381.84722 - Fax 0381.73393 - E-mail: [info@cimaonline.com](mailto:info@cimaonline.com) - Internet: <http://www.cimaonline.com>

According to the EEC instruction 89/686 dated 21st of December 1989, concerning the standardisation of legislation of all Member Countries with regards to individual protection system and the relative legislative decree dated 4th of December 1992, N°. 475,

**A.N.C.I. Servizi s.r.l. a socio unico - C.I.M.A.C. section  
CENTRO ITALIANO MATERIALI DI APPLICAZIONE CALZATURIERA  
Authorized with the decree issued by the Ministry of Industry of the Italian Republic  
On the 11<sup>th</sup> of October 2000 – Community identification N°. 0465**

grants:

## **CLOSURE REPORT OF EC CERTIFICATION**

**N° 0162/24024/16 - 00863C**

### **MODULE C2 – EC TYPE CONFORMITY BASED ON INTERNAL PRODUCTION CONTROL WITH TESTS MADE UNDER CONTROL OFFICER AT RANDOM INTERVALS**

For the following model of personal protective equipment of Cat. III:

**Protective glove against chemicals and micro organism - Article  
"FLEXISKIN bluple nitrile glove 240 mm"**

Manufacturer (see notes):

**LABORATOIRES EUROMEDIS**

**ZA DE LA TUILERIE  
60290 NEUILLY SOUS CLERMONT**

**The validity of the closure report shall be subject to surveillance activities provided for by Directive 89/686/EEC, to be implemented within 12 months from the issue date specified below.**

Vigevano, 23 May 2017

Responsible of CE Certification of Glove  
Dr. Sandro Milanese

The Technical Responsible  
Ing. Giuseppe Bellotti



## 1. Description of personal protective equipment:

Category of PPE: third category

Attestato di certificazione CE del tipo 0162/24024/16

Type of PPE: Protective glove against chemicals and micro organism

In compliance with EN 374: 2003 Part I II and III

Protection against anti neoplastic agents

Model of glove: glove with five fingers

Size range: from 6 to 10 (from XS to XL)

Manufacturing process: dipping





**2. The tests and the examinations to verify the conformity of the article (in compliance with art. 11A of Directive 89/686/EEC – Decision 768/08/EC Module C2) are performed applying the following harmonized standards:**

Standards applied:

- A - EN 420:2003 + A1: 2009 – General requirements for gloves.
- B - EN 374:2003 part I II e III – Glove against chemicals and micro organism
- C - EN 388:2016 – Protective gloves against mechanical risk.

**3. The results of tests and examinations are contained in the following test reports:**

C.I.M.A.C.    2017/0936-RP -1-RP-1                          Dated                          23 of May                          2017

**4. Requirements of the personal protective equipment:**

Based on the test carried out, the model of PPE examined is conform to the model described in the CE type certificate and that there is uniformity in production activities.

The model of protective glove against chemicals and micro organism - Article "FLEXISKIN bluple nitrile glove 240 mm" conforms:

- the requirements of EN 420:2003 + A1 2009 standard, points 4.1, 4.2, 4.4, 5 and 5.2;
- the following permeation levels as specified in EN 374:2003 Part I standard;
- the following permeation performance levels as specified in EN 374:2003 Part I;

	Performance level
Sulphuric acid solution at 96%	2
Sodium hydroxide solution at 40%	3
Dyethylamine	2

- the performance levels as specified in prospectus 1 of EN 388:2016 standard;

	Performance level
6.1 Abrasion resistance	0
6.2 Cut resistance	0
6.3 Tear resistance	0
6.4 Puncture resistance	0
5.2 Dexterity	5

In the model of Personal Protective Equipment and its components was not detected the presence of dangerous substances listed in Annex XVII of Regulation 1907/2006/EC and subsequent amendments and additions.



## 5. Marking of the personal protective equipment:

The following information is provided on the box of the glove:

- the "CE" mark
- the article code: FLEXISKIN bluple nitrile glove 240 mm
- the manufacturer name: LABORATOIRES EUROMEDIS
- the glove size
- the symbols concerning the protection provided: Protective gloves against chemicals and micro organisms.
- the permeation performance levels as specified in EN 374:2003 Part I standard:

	Performance level	Code
Sulphuric acid solution 96%	2	L
Sodium Hydroxide solution 40%	3	K
Dyethylamine	2	G

- the performance levels as specified in prospectus 1 of EN 388:2016 standard;

	Performance level
6.1 Abrasion resistance	0
6.2 Cut resistance	0
6.3 Tear resistance	0
6.4 Puncture resistance	0
5.2 Dexterity	5



## 6. Notes:

- "*making available on the market*" shall mean any supply of a product for distribution, consumption or use on the Community market in the course of a commercial activity, whether in return for payment or free of charge.
- "*placing on the market*" shall mean the first making available of a product on the Community market.
- "*manufacturer*" shall mean any natural or legal person who manufactures a product or has a product designed or manufactured, and markets that product under his name or trademark.
- "*authorised representative*" shall mean any natural or legal person established within the Community who has received a written mandate from a manufacturer to act on his behalf in relation to specified tasks.
- "*harmonised standard*" shall mean a standard adopted by one of the European standardisation bodies listed in Annex I to Directive 98/34/EC on the basis of a request made by the Commission in accordance with Article 6 of that Directive.
- "*accreditation*" shall mean an attestation by a national accreditation body that a conformity assessment body meets the requirements set by harmonised standards and, where applicable, any additional requirements including those set out in relevant sectoral schemes, to carry out a specific conformity assessment activity.
- "*conformity assessment*" shall mean the process demonstrating whether specified requirements relating to a product, process, service, system, person or body have been fulfilled.
- "*recall*" shall mean any measure aimed at achieving the return of a product that has already been made available to the end user.
- "*withdrawal*" shall mean any measure aimed at preventing a product in the supply chain from being made available on the market.
- "**CE** marking" shall mean a marking by which the manufacturer indicates that the product is in conformity with the applicable requirements set out in Community harmonisation legislation providing for its affixing.
- The content of this EC Type-Examination Certificate is referred to the tested personal protective equipment only.
- This EC Type-Examination Certificate may be integrally duplicated; the copy must be faithful, legible (if pint size) and must contain the bold caption "TRUE COPY".



**A.N.C.I. Servizi S.r.l.**  
a socio unico

Sede legale e amministrativa  
20149 MILANO  
Via Monte Rosa, 21  
Tel. 02.438291  
Fax 02. 48005833

Cap. Soc.: € 10.400 i.v.  
C.F./P.I.: 07199040150  
Reg. Imprese n° 229059  
Trib. di Milano  
R.E.A. n° 1147818

Sede operativa: 27029 VIGEVANO (PV) - C.so G. Brodolini, 19 - Tel. 0381.84722 - Fax 0381.73393 - E-mail: info@cimaonline.com - Internet: http://www.cimaonline.com



**CIMAC**  
Centro Italiano  
Materiali di Applicazione  
Calzaturiera



LAB N° 0005

## LABORATOIRES EUROMEDIS

**ZA de la TUILERIE**  
**NEUILLY sous CLERMONT**  
**60290**

**TEST REPORT: RP 2017/0936-1-RP-1 of 23/05/2017**

**Page 1 of 4**

SAMPLE RECEIVED ON: 10/05/2017

SAMPLE RECEIVED FOR TESTING:

Protective glove against chemicals and micro organism - Article "FLEXISKIN bluple nitrile glove 240 mm"

TEST REQUESTED:

Determination of the characteristics according with EN 374:2003 Part I II e III standard  
Determination of the characteristics according with EN 388: 2016

		
EMISSIONE	P. BIGLIA	S. MILANESI
OGGETTO	<b>RESPONSABILE LAB. FISICO MECCANICO</b>	<b>RESPONSABILE LAB. ANALISI CHIMICHE</b>

Il campionamento del materiale ricevuto da esaminare, se non diversamente indicato, è stato effettuato dal cliente.

Il residuo del campione analizzato si conserva per tre mesi.

Il Rapporto di Prova non ha validità di approvazione e/o certificazione del campione esaminato.

Il marchio ACCREDIA e/o l'Accreditamento del CIMAC non possono essere utilizzati nella documentazione di prodotto a meno che non venga riportata copia integrale, fedele, leggibile del rapporto di prova contenente la dicitura in grassetto "Copia Conforme all'Originale".

Il CIMAC è accreditato da ACCREDIA con il numero di accreditamento 0005. Per le prove accreditate il ACCREDIA garantisce la competenza del personale, la disponibilità di strumentazione e la conformità delle procedure di prova alla norma/procedura richiamata.

Il contenuto del presente Rapporto di Prova si riferisce unicamente al campione sottoposto a prova.

Le prove riportate nel presente Rapporto di Prova contrassegnate dalla dicitura " Non accreditate da ACCREDIA" non rientrano nell'Accreditamento.



LAB N° 0005



Physics-mechanical laboratory and chemical analysis

Tests carried out from 11.05.17 to 23.05.17

No. 500 samples of protective gloves against chemicals and micro organism sampled by CIMAC at the storage center of Laboratoires Euromedis – Z.A de la Tuilerie – 60290 Neuilly sous Clermont on 09.05.2017

**Determination of the characteristics according to EN 374:2003 Part I II and III standard.**

**Determination of the characteristics according to EN 388:2016**

**Periodic control of PPE – Batch number: SU1609976**

Reference to test register: from F/5419 to F/5421

**Items of EN 420:2003 + A 1 2009**

**4 General**

**4.2 Construction**

The material and the construction are such that do not decrease the general performances of the glove.

**4.4 Innocuousness of protective gloves**

**Palm - Back**

(Nitril - cod. Bluple nitrile)

**5 Comfort and efficiency**

**5.1 Sizes**

**5.1.2 Sizes and measures of gloves**

Length of the glove

size 6 left = 240 mm

size 6 right = 220 mm

size 8 left = 250 mm

size 8 right = 250 mm

size 11 left = 260 mm

size 11 right = 260 mm

**5.2 Dexterity**

The smallest diameter of pin fulfilling test condition is 5 mm.

**Item of EN 374:2003 Part II**

**5.2 Air leak test**

6 The glove pass the test



LAB N° 0005

**5.3 Water leak test**

6 The glove pass the test

**Item of EN 374:2003 Part III****8.4 Assessment of breakthrough time**

8.5.1 Tested chemical product: Sulphuric acid

Révélation technique: potentiometrical analyses

Concentration of test solution : 96% by weight

Test temperature:  $23 \pm 2^{\circ}\text{C}$ 

Sample from the palm of the glove

Average thickness of the sample = 0.08 mm

Collection agent: water.

Sample appearance after testing: colour variation – turn to brown – sample brittle

Results:

After 30' it is not verified the passage of the tested chemical product Sulphuric acid in the collection agent in quantity over  $1\mu\text{g min}^{-1}\text{ cm}^{-2}$  (passage occur at 35')**Items of EN 388:2016 (\*)**

Results:

**Palm – Back**

(Nitril - cod. Bluple nitrile)

**4.0 Requirements**

Table "1"

**6.1 Abrasion resistance**

Before 100 test cycles the surface of the specimens is broken.

**6.2 Blade cut resistance**

Sequence	Check specimen C	Specimen T	Check specimen C	Index $I_n$
1A	2	0	2	1.0
2A	2	0	2	1.0
3A	2	0	2	1.0
4A	2	0	2	1.0
5A	2	0	2	1.0

Sequence	Check specimen C	Specimen T	Check specimen C	Index $I_n$
1B	2	0	2	1.0
2B	2	0	2	1.0
3B	2	0	2	1.0
4B	2	0	2	1.0
5B	2	0	2	1.0

Minimum index = 1.5



LAB N° 0005



**6.3 Tear resistance**

Specimens taken in direction of the length of the palm of the glove	8 N
	9 N
Specimens taken in direction of the width of the palm of the glove	6 N
	5 N

**6.4 Puncture resistance**

Specimen "1"	7 N
Specimen "2"	6 N
Specimen "3"	5 N
Specimen "4"	6 N

(\*) – test not accredited by ACCREDIA

\* End of Test Report \*

---