



Barrier Product Solutions

# STERIS Barrier Product Solutions

## CERTIFICATE OF CONFORMANCE

Date: December 2017

Company: HLD CLEAN CONSULT

To: Q.C. Department

P.O. #: C-125394

Reference #: N/A

Item Reference #: N/A

GEPCO Re-Order #: 1805TWSTB

Product Description: SLEEVE/ARMCOVER, WHITE, 1 PAIR/PACK,  
50 PACKS/POLYBAG, 5 POLYBAGS PER  
POLYLINE CASE, STERILE

Batch Control #: 127C/20,20

Date of Manufacture: 12/17

Expiration Date: 12/20

We certify that this product was manufactured in accordance to all described specifications, as indicated on the STERIS BPS PRODUCT DATA SHEET.

Signature:

Quality Assurance Representative

Rev. Dec 2016

CERTIFICATE OF STERILIZATION

Sterilization Lot No.: 127C120.20

Sterilization Date: 12/5/17

18 biological indicators were placed into the sterilization containers at various locations for monitoring of the sterilization process.

Products sterilized within the above designated lot number, were monitored for sterility utilizing Mesa Labs "EZTEST" Biological indicators.

After 48 hours incubation of the EZTEST monitors, the results were as follows:

<u>BIOLOGICAL INDICATORS:</u>	<u>LOT NO. (s)</u>	<u>RESULTS</u>
<i>Bacillus atrophaeus</i>	<u>6204</u>	<u>0-18</u>

This sterilization lot meets the requirements of the test for sterility.

Christopher Thompson  
Production Department Representative

12/11/17  
Date

Released by: [Signature]  
Quality Assurance Representative

11 Dec 2017  
Date

Rev. Sep 2016